

## Handling of EWL, EAL and EGL series samplers

EWL, EAL and EGL series samplers are delivered packed by one in air tight can. Air tight can should be open only in the place of deployment to avoid contamination of membrane.



To open the can is very practical to use bigger coin. It should be opened carefully moving coin around in order not to damage lid.



The same can could be used for the retrieved membranes, if the lid was not damaged during opening. Close can with all hand that it would be even pressure around. Press it till it would click.

